


**Customer** ESS Data Recovery  
**Address** 110 North Research Dr  
**City/State** Edwardsville, IL 62025  
**Department** n/a  
**Phone** (800) 237-4200

**Make:** Envirco  
**Model:** Z04  
**Serial Num.** S0P-120V-00790

Location	Counter	Status	TimeStamp	Period	0.3	0.5	1	5	FLO	CRC
1	Met One	OK	10/9/2012 9:03	0:01:00	33	2	0	0	95	OK
2	Met One	OK	10/9/2012 9:04	0:01:00	29	1	0	0	96	OK
3	Met One	OK	10/9/2012 9:05	0:01:00	40	0	0	0	95	OK
4	Met One	OK	10/9/2012 9:06	0:01:00	36	1	0	0	96	OK
5	Met One	OK	10/9/2012 9:07	0:01:00	39	2	0	0	96	OK



Date: 10/9/12

Field Service Engineer *Joseph Burch*

# Federal Standard 209E Report

using PortAll V2.0

Report Date: 10/9/2012

**ACE ELECTRIC LAB SYSTEMS**  
 1550 KINGSHIGHWAY  
 ST. LOUIS , MO 63110

Room or Area:	S09-120V-00790	Start Date/Time:	10/9/2012 9:00 AM
Size (sq M):	1.00	Finish Date/Time:	10/9/2012 9:12 AM
Target Class:	100	Result:	Passed

**Summary Data: Calculations are based on cubic meters**

Sizes	0.3µm	0.5µm
Mean	1,250.14	42.38
Standard Deviation	159.11	29.55
Standard Error	71.16	13.21
95% UCL	1,401.70	70.52
Class 100 Limits	10,600.00	3,530.00
Pass/Fail	Pass	Pass

**Locations sampled:**

1	2	3	4	5
---	---	---	---	---

**Instruments:**

Name:	Met One	Serial Number:	000100515
-------	---------	----------------	-----------

# ISO 14644-1 Report

using PortAll V2.0

Report Date: 10/9/2012

ACE ELECTRIC LAB SYSTEMS  
1550 KINGSHIGHWAY  
ST. LOUIS , MO. 63110

Room or Area:	S09-120V-00790	Start Date/Time:	10/9/2012 9:00 AM
Size (sq M):	1.00	Finish Date/Time:	10/9/2012 9:13 AM
Target Class:	5	Result:	Passed

## Summary Data:

Sizes	0.3µm	0.5µm
Mean	1,250.14	42.38
Standard Deviation	159.11	29.55
Standard Error	71.16	13.21
95% UCL	1,399.57	70.13
Class 5 Limits	10,200.00	3,520.00
Pass/Fail	Pass	Pass

## Locations sampled:

1	2
3	4
5	

## Instruments:

Name: Met One	Serial No.: 000100515
---------------	-----------------------